

Calibration of Peltier Heat Flux Plate (HFP) serial number P03
University of British Columbia, Jan 7 - 16, 2013

Reference Plate:

Hukseflux HFP01-15

SN: 6649

Multiplier: 62.0 $\mu\text{V}/(\text{W}/\text{m}^2)$
16.13 $(\text{W}/\text{m}^2)/\text{mV}$

Test Plate:

TE Technology HP-127-1.0-1.3-71P

SN: P03 (assigned by biomet lab)

Multiplier: 100.2 $(\text{W}/\text{m}^2)/\text{mV}$

Note: linear regression forced through zero

