



**Hukseflux Thermal Sensors B.V.**  
URL: [www.hukseflux.com](http://www.hukseflux.com)  
E-mail: [info@hukseflux.com](mailto:info@hukseflux.com)  
The Netherlands

## Certificate of calibration

Type: HFP01-15  
Serial number: 10867

Received 2015-09-28

Calibration date: 10-11-2014  
Cable length: 15 m  
Sensitivity: 62.17  $\mu\text{V}/(\text{W}/\text{m}^2)$   
Remarks:  
Performed by: Y. Putter

### Connections:

Cable	Wire	
1	White	Signal [+]
1	Green	Signal [-]
1	Black	Shield

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.