

Certificate of calibration

Type: HFP01-15
Serial number: 11329

Received 2015-09-28

Calibration date: 11-06-2015
Cable length: 15 m
Sensitivity: 61.47 $\mu\text{V}/(\text{W}/\text{m}^2)$
Remarks:
Performed by: Y. Putter

Connections:

Cable	Wire	
1	White	Signal [+]
1	Green	Signal [-]
1	Black	Shield

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.