



Hukseflux Thermal Sensors B.V.
URL: www.hukseflux.com
E-mail: info@hukseflux.com
The Netherlands

Certificate of calibration

Type: HFP01-15
Serial number: 11334

Received 2015-09-28

Calibration date: 11-06-2015
Cable length: 15 m
Sensitivity: 61.04 $\mu\text{V}/(\text{W}/\text{m}^2)$
Remarks:
Performed by: Y. Putter

Connections:

| Cable | Wire | |
|-------|-------|------------|
| 1 | White | Signal [+] |
| 1 | Green | Signal [-] |
| 1 | Black | Shield |

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.