

Hukseflux Thermal Sensors B.V.

www.hukseflux.com info@hukseflux.com

Product certificate

Pages:

1

Release date:

10-04-2015

Product code

NR01-00

Product identification

serial number 2420

Product type Measurand 4-component net-radiation sensor

net radiation

Product specifications

functional test
cable length

passed

0 m

Person authorising acceptance and release of product: W.J.B. Fokke

Date:

10-04-2015

Calibration results

Component	SR01↓	SR01 ↑	IR01 J	IR01↑
Position	1 2 3 3 3 2 3 3 3 3 3 3 3	2	3	4
Serial number	4045	4046	3986	3987
Sensitivity	19.56 x 10 ⁻⁶	16.38 x 10 ⁻⁶	11.50 x 10 ⁻⁶	13.15 x 10 ⁻⁶
Uncertainty*	± 0.24 × 10 ⁻⁶	± 0.20 x 10 ⁻⁶	± 0.61 x 10 ⁻⁶	± 0.70 x 10 ⁻⁶
Calibration date	30-01-2015	30-01-2015	04-02-2015	03-02-2015
Resistance	66.7	63.5	169.2	204.1

^{*} the number following the \pm symbol is the expanded uncertainty with a coverage factor k=2, and defines an interval estimated to have a level of confidence of 95 percent

Table 0.1 connections Cable 1

Table 0.2 connections Cable 2

PCB04	WIRE		PCB05	WIRE	
1	Blue	SR01 ↓ [-]	1	Вгомп	heater
2	Red	SR01 ↓ [+]	2	APS	Pt100 [+]
3	Yellow	IR01 ↓ [-]	3	White	Pt100 [+]
4	Brown	IRO1 ↓ [+]	4	Blue	Pt100 [-]
5	Grey	IR01 ↑ []	5	Green	Pt100 [-]
6	Pink	IR01 ↑ [+]	6	Yellow	heater
7	Green	SR01 ↑ [-]		Pink	not connected
8	White	SR01↑[+]		Grey	not connected
11	Black	shield	10	Black	shield

The Pt100 is a single four-wire resistance temperature detector measuring instrument body temperature.

SR01 calibration procedure according to ISO 9847. Traceability of calibration is to the WRR (World Radiometric Reference) maintained at the World Radiation Center in Davos, Switzerland.

IR01 calibration procedure according to Hukseflux IRC02. Traceability of calibration is to the WISG (World Infrared Standard Group) operated at the Infrared Radiometry Section of the World Radiation Center in Davos, Switzerland.

Please consult the user manual for detailed measurement functions and product set up, operation and maintenance instructions.