

Campbell Scientific Limited

80 Hathern Road,
Shepshed,
Leicestershire,
LE12 9GX, United Kingdom.



Telephone: +44 (0) 1509 601141

HygroVUE™10 Calibration Check

Serial Number	Reference RH (%)	Probe RH (%)	Reference Temperature (°C)	Probe Temperature (°C)
E3514	33.60	31.35	22.59	22.61

The Test Standards used have full traceability to UK Standards. All standards have been calibrated by Campbell Scientific or a UKAS accredited calibration service.

Check performed by: Keri

Date: 29th March 2022.

Campbell Scientific Limited operates a Quality Assurance System approved to BS EN ISO 9001:2015