



**CAMPBELL  
SCIENTIFIC**<sup>®</sup>

WHEN MEASUREMENTS MATTER

## *Certificate of Conformance*

Campbell Scientific, Inc. hereby certifies that the SoilVUE10 TDR Soil Moisture, Electrical Conductivity and Temperature Profile Sensor is assembled at the Logan, UT facility and is calibrated in dielectric and NIST traceable conductivity standard solutions using laboratory reference instrumentation and is inspected and tested to meet Campbell Scientific published specifications. The SoilVUE10 is also inspected to the IPC-A-610G Acceptability of Electronic Assemblies standard.

Campbell Scientific, Inc is a registered ISO 9001:2015 company, and ensures compliance to relative regulatory standards.

Reference Standard: TDR200 SN 1345 and CS547A conductivity sensor SN 7487 and 7489.

SoilVUE10 SN: *2676*

Date: 30 Nov 2020

By:

*Quality Assurance*

