



**Hukseflux Thermal Sensors B.V.**

URL: [www.hukseflux.com](http://www.hukseflux.com)

E-mail: [info@hukseflux.com](mailto:info@hukseflux.com)

The Netherlands

## Certificate of calibration

Type: HFP01-15

Serial number: 6492

Calibration Date: 22-12-2011

Sensitivity: 61.3  $\mu\text{V}/\text{Wm}^{-2}$

Remarks:

Performed By: Chris VAN SOEST

Connections

White thermopile +

Green thermopile -

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.