



Hukseflux Thermal Sensors B.V.

URL: www.hukseflux.com

E-mail: info@hukseflux.com

The Netherlands

Certificate of calibration

Type: HFP01-15

Serial number: 6499

Calibration Date: 22-12-2011

Sensitivity: 62.6 $\mu\text{V}/\text{Wm}^{-2}$

Remarks:

Performed By: Chris VAN SOEST

Connections

White thermopile +

Green thermopile -

Traceability of calibration is to the "guarded hot plate" of National Physical Laboratory (NPL) of the UK. Applicable standards are ISO 8302 (EN 12664:2001) and ASTM C177.