



Kipp & Zonen B.V.

Röntgenweg 1, 2624 BD Delft

P.O. Box 507, 2600 AM Delft

The Netherlands Tel: +31 15 269 8000 Fax: +31 15 262 0351

Canada. France. Germany. United Kingdom. USA.

old

Calibration certificate

Instrument Type : CNR1

Serial No : 990181

Sensitivity : 10.19 Microvolt/Watt/Metre square

Options : none

Cable length : 10 Meters

Connectors : none

Reference conditions : 20 degrees Celcius
horizontal position
500 Watt/Metre square

Radiation Scale : World Radiometric Reference

Reference for CM3 : CM3 FT014

Reference for CG3 : CG1 FT FT002

Calibration date : 20/08/99

Calibrated by : G.v.d.Wilt



Kipp & Zonen B.V.
Delft Holland