



**CALIBRATION CERTIFICATE**

*Recd 21 Aug 01*

**QUANTUM SENSOR**

Calibration date : 26-juli-2001

Calibrated by : G. Lindner

Quantum sensor, type : **PAR LITE**

Serial no : 010163

Sensitivity ( $\pm 2\%$ ) : 5.32  $\mu V/\mu mol/s.m^2$   
in solar radiation at airmass 1.5

Impedance : 240 Ohm

Calibration procedure : Exact interchange of test PAR LITE and reference PAR LITE in a horizontal parallel beam from a Xenonlamp. Photosynthetic photon flux density approx. 400  $\mu mol/s.m^2$ . Instrument temperature approx. 25° C.

Hierarchy of traceability : The reference PAR LITE is calibrated against a standard of known illuminance, the photometric standard lamp of Osram type Wi 41/G with color temperature of 2856 K, which on his turn is calibrated yearly at the dutch standard laboratory NMI. From the known illuminance and the 2856 K spectrum the photosynthetic photon flux density is calculated: 6.122  $\mu mol/s.m^2$ .  
Correction applied : The sensitivity figure of the reference PAR LITE is made correct for hemispherical solar irradiance at airmass 1.5 (spectrum according international standard ISO 9845-1). The correction of the sensitivity figure was +1.2 %.

