

Recd 10 Jul 04 TY



**Kipp &
Zonen**

Kipp & Zonen B.V.
Röntgenweg 1 2624 BD Delft
P.O. Box 507 2600 AM Delft
The Netherlands
T +31-(0)15-2698000
F +31-(0)15-2620351
E info.holland@kippzonen.com
Website www.kippzonen.com

CALIBRATION CERTIFICATE

QUANTUM SENSOR

Calibration date : 29-05-04

Calibrated by : G.Lindner

Quantum sensor, type : **PAR LITE**

Serial no : 040411

Sensitivity ($\pm 2\%$) : 5.66 $\mu\text{V}/\mu\text{mol}/\text{s}.\text{m}^2$
in solar radiation at airmass 1.5

Impedance : 240 Ohm

Calibration procedure : Exact interchange of test PAR LITE and reference PAR LITE in a horizontal parallel beam from a Xenonlamp. Photosynthetic photon flux density approx. 400 $\mu\text{mol}/\text{s}.\text{m}^2$. Instrument temperature approx. 25^o C.

Hierarchy of traceability : The reference PAR LITE is calibrated against a standard of known illuminance, the photometric standard lamp of Osram type Wi 41/G with color temperature of 2856 K, which on his turn is calibrated yearly at the dutch standard laboratory NMI. From the known illuminance and the 2856 K spectrum the photosynthetic photon flux density is calculated: 6.122 $\mu\text{mol}/\text{s}.\text{m}^2$. The sensitivity figure of the reference PAR LITE is made correct for hemispherical solar irradiance at airmass 1.5 (spectrum according international standard ISO 9845-1). The correction of the Correction applied to sensitivity figure was +1.2 %.

File name: PARLITEcalibratie cert..doc
Date : 980701