



**Kipp & Zonen B.V.**

Röntgenweg 1, 2624 BD Delft

P.O. Box 507, 2600 AM Delft

The Netherlands Tel: +31 15 269 8000 Fax: +31 15 262 0351

Canada. France. Germany. United Kingdom. USA.

## CALIBRATION CERTIFICATE

### QUANTUM SENSOR

Calibration date : 18-Jun-99

Calibrated by : F. de Wit

Quantum sensor, type : **PAR LITE**

Serial no : 990025

Sensitivity ( $\pm 2\%$ ) : 5.03  $\mu\text{V}/\mu\text{mol}/\text{s}\cdot\text{m}^2$   
in solar radiation at airmass 1.5

Impedance : 240 Ohm

Calibration procedure : Exact interchange of test PAR LITE and reference PAR LITE in a horizontal parallel beam from a Xenonlamp. Photosynthetic photon flux density approx. 400  $\mu\text{mol}/\text{s}\cdot\text{m}^2$ . Instrument temperature approx. 25° C.

Hierarchy of traceability : The reference PAR LITE is calibrated against a standard of known illuminance, the photometric standard lamp of Osram type Wi 41/G with color temperature of 2856 K, which on his turn is calibrated yearly at the dutch standard laboratory NMI. From the known illuminance and the 2856 K spectrum the photosynthetic photon flux density is calculated: 6.122  $\mu\text{mol}/\text{s}\cdot\text{m}^2$ .

Correction applied : The sensitivity figure of the reference PAR LITE is made correct for hemispherical solar irradiance at airmass 1.5 (spectrum according to international standard ISO 9845-1). The correction of the sensitivity figure was +1.2 %.