

CERTIFICATE of CALIBRATION for LI-COR SENSOR

Quantum Sensor
Model Number: LI-190/R

Received 2015-06-11
Serial Number: Q100421

Calibration Date: May 08, 2015
Manufacture Date: May 01, 2015

Calibration Constant:

Output: 6.63 microamps per 1000 $\mu\text{mol s}^{-1} \text{m}^{-2}$

For use with LI-COR handheld meters and loggers:

Multiplier: -150.92 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per microamp

For use with LI-190R-BL (3-wire bare leads):

Multiplier: 150.92 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per microamp

For use with LI-COR 2290 (604 ohm) Millivolt Adapter:

Multiplier: -249.86 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per millivolt

If this is an SL or SMV sensor:

Multiplier: -200.0 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per millivolt

IMPORTANT: Read the appropriate instruction manual (www.licor.com/Quantum-manual) before using this sensor.

IMPORTANT: It is recommended that sensors be recalibrated every two years after field deployment.

Calibration Technician: *Cameron Oedema*

Calibration standard used: Working standard lamp 811D.
Calibration traceable to the National Institute of Standards and Technologies (NIST) through NIST-calibration lamp number F-226.
(NBS Test No. 534/237689-86)

LI-COR

LI-COR Biosciences, Global Headquarters • 4647 Superior Street • Lincoln, NE 68504 USA
Phone: +1-402-467-3576 • Fax: +1-402-467-2819 • Toll-free: 800-447-3576 (USA & Canada)
envsales@licor.com • envsupport@licor.com • www.licor.com
Manual: www.licor.com/Quantum-manual