

LI-7810 CH₄/CO₂/H₂O Analyzer

Calibration Certificate

Serial Number TG10-01751

Date: 2023-07-31

Technician: 
J.D.S.

Verification Gas

Assigned Values

2010.5000 ppb CH₄

3995.1000 ppb CH₄

405.4000 ppm CO₂

1006.0000 ppm CO₂

Unit Under Test

Measurements

2010.4692 ppb CH₄

4011.5026 ppb CH₄

406.9800 ppm CO₂

1010.3230 ppm CO₂