



Precision Measurement Engineering, Inc.  
1487 Poinsettia Ave. Suite 129  
Vista, CA 92081

Telephone (760) 727-0300

**Specification/ Calibration Certificate**

Product: miniDOT Logger  
Application: Dissolved oxygen and Temperature

This certificate confirms that the miniDOT Logger with **Serial Number(s)**

899502, 906606, 931528

has passed all manufacturing performance testing.

The certificate confirms that this logger has undergone the proprietary detailed and intensive calibration process to ensure that it meets the specifications below.

Date of calibration was performed on March 2018 by E. Elliott.

For operating instructions, please refer to the miniDOT Logger User's Manual, which can be found on your miniDOT logger once connected to a computer. Or it can be downloaded from the PME web site here,

<http://pme.com/products/minidot>.

Temperature Accuracy	+/-0.1 degrees C
Temperature Range	0 to 35 degrees C
Oxygen Accuracy	+/-10 umole/L
Oxygen Range	0 to 150% saturation

Thank you for your purchase! Please email us at [info@pme.com](mailto:info@pme.com) if you have any questions.

Best Regards,

PME Team