



December 16,

2016

Specification Certificate

Product: C3 Submersible Fluorometer

This certificate confirms that the C3 Submersible Fluorometer **Serial Number: 2300565** has passed all manufacturing performance testing and meets the specifications detailed below.

For operating instructions, please refer to the C3 and C6P Submersible Fluorometer User's Manual contained on the USB flash drive that is included with the unit or can be viewed at the following web link:

<http://www.turnerdesigns.com/t2/doc/manuals/998-2300.pdf>

Measurement Parameter	Channel 1	Channel 2	Channel 3
Sensor	Chlorophyll a	Turbidity	CDOM
Blank Test (+/- 5% of 0.00 mV to 35.00mV)	Passed	Passed	Passed
Noise Test (< 5 mV)	Passed	Passed	Passed
Gain Function (1X – 100X)	Passed	Passed	Passed