

CALIBRATION CERTIFICATE

Instrument GMP343 Carbon dioxide probe
Serial number N2220007
Manufacturer Vaisala Oyj, Finland
Calibration date 31st May 2017

The carbon dioxide probe was calibrated against accurate bottled nitrogen and carbon dioxide gas concentrations. High purity nitrogen gas was used for calibrating the zero value of the probe. For the CO₂ calibration, an accurate mixture of nitrogen and carbon dioxide traceable to VSL (Dutch Metrology Institute) was used.

Pressure and temperature compensation was made by using the compensation algorithm of the probe. The ambient pressure value was obtained from a Vaisala pressure transmitter, and the gas temperature was measured by the built-in temperature sensor of the probe. The temperature sensor of the probe was calibrated against the gas temperature sensor of the calibration system. At the time of shipment, the probe met its operating specifications.

The pressure transmitter and the gas temperature sensors have been calibrated at ISO/IEC 17025 accredited calibration laboratory (FINAS), MSL by using working standards traceable to the National Institute of Standards and Technology (NIST).

Calibration results (for measurement range 0...1000 ppm)

Reference ppm CO ₂	Observed* ppm CO ₂	Difference ppm CO ₂	Permissible difference ppm CO ₂	Calibration gas temperature °C	Ambient pressure hPa
0	1	1	± 3	26.5	1004.6
200	201	1	± 5	26.5	1004.7
371	373	2	± 6	26.5	1004.9
600	602	2	± 9	26.5	1005.0
1000	1006	6	± 13	26.5	1005.1

*Reading after pressure and temperature compensation.

Ambient conditions / Humidity 17 ± 5 %RH, Temperature 25.0 ± 1 °C.

Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
34970A	MY41021619	2017-03-15	1250-307082540
PT-100 Sensor	TP12074	2017-05-09	A00974

Gas cylinders used in calibration

Type	CO ₂ concentration (ppm)	Cylinder number	Reference number
Scientific nitrogen (N6.0)	0	7525010202439	100396693
Carbon dioxide mixture (± 0.5 %)	200	7520080002415	100443691
Carbon dioxide mixture (± 0.5 %)	371	7523112125581	100458390
Carbon dioxide mixture (± 0.5 %)	600	7523112099553	100458931
Carbon dioxide mixture (± 0.5 %)	1000	7523610208596	100430081



Technician

This report shall not be reproduced except in full, without the written approval of Vaisala.

DOC214909-D