

## CALIBRATION CERTIFICATE

after adjustment

**Customer** Campbell Scientific  
**Instrument** PTB101B Analog barometer  
**Serial number** B0310006  
**Manufacturer** Vaisala Oyj, Finland  
**Calibration date** 8th July 2015  
**Test procedure** doc210609a

This instrument has been calibrated against a Vaisala PTB220 factory working standard which has been calibrated against a Ruska 2465 pressure balance traceable to the National Institute of Standards and Technology (NIST, USA) at Vaisala Measurement Standards Laboratory (MSL). Vaisala MSL has been accredited by the Finnish Accreditation System (FINAS) according to ISO/IEC 17025 standard.

At the time of shipment, the instrument described above met its operating specifications.

### Calibration results

Reference pressure hPa	Observed pressure hPa	Correction* hPa	Uncertainty** hPa
619.6	619.6	0.0	± 0.15
699.6	699.6	0.0	± 0.15
799.7	799.6	0.1	± 0.15
849.6	849.5	0.1	± 0.15
899.6	899.5	0.1	± 0.15
949.6	949.5	0.1	± 0.15
999.6	999.5	0.1	± 0.15
1059.7	1059.8	-0.1	± 0.15


\*To obtain the true pressure, add the correction to the barometer reading. Interpolated corrections may be used at intermediate readings of the scale of the barometer.

\*\*The calibration uncertainty given at 95 % confidence level, k = 2

### Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
Vaisala PTB220	X1260001	2015-01-29	150129-PTB220- X1260001
Vaisala PTB220	X3710015	2015-01-29	150129-PTB220- X3710015
34970A	MY44052432	2014-08-21	516173

**Ambient conditions** / Humidity  $48 \pm 5$  %RH, Temperature  $23 \pm 1$  °C, Pressure  $1009 \pm 1$  hPa

  
 Matthew Nocivelli

*This report shall not be reproduced except in full, without the written approval of Vaisala.*

doc210635c