



Biosciences

LI-COR Service Report

RMA 00186965

RMA Received on: 01/13/2015

Customer Information:

University of California - Berkeley
BERKELEY, CA 94720-3114
UNITED STATES

Joseph Verfaillie
jverfail@berkeley.edu

Description of Service						
<u>Serial #</u>	<u>Part</u>	<u>Qty</u>	<u>Parts Total</u>	<u>Labor Hrs.</u>	<u>Labor Total</u>	<u>Line Item Subtotal</u>
AIU-0397	9672-009 : CONTR 7550 ANALOG PCA	1	\$ 1325.00			\$ 1325.00
*TOTAL:						\$1,325.00

**Total does NOT include shipping charges.*

Notes on Charges:

RMA Model & Serial Number(s)	
<u>Model</u>	<u>Serial Number</u>
LI-7550	AIU-0397

Service Notes
 Inspected the unit and found the LI-7550 would not power up. Found multiple power supplies shorted due to ants getting inside the LI-7550. Removed ants, replaced the defective analog board and subjected the unit to environmental tests.